

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00031
Calibration File No.:	00785	Calib. Date:	08/20/2015
Certification File No.:	00786	Cert. Date:	08/20/2015
Linearity File No.:	00787	Lin. Date:	08/20/2015
Solution File No.:	00830	Soln. Date:	02/09/2016
Sequential File No.:	00830	File Date:	02/09/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	0082

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:07S		
Control 1 EC	0.100%	12:08S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:08S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		
Control 2 EC	0.100%	12:09S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:09S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10S		
Control 3 EC	0.100%	12:11S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:11S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: 

Badge No.: 421

Date: 02/09/2016