Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C		Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00705	Calib. Date: 09/30/2014	Calib. No.: 00029
Certification File No.:	00706	Cert. Date: 09/30/2014	Cert. No.: 00019
Linearity File No.:	00707	Lin. Date: 09/30/2014	Lin. No.: 00019
Solution File No.:	00738	Soln. Date: 02/16/2015	Soln. No.: 00160
Sequential File No.:	00738	File Date: 02/16/2015	

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %:	0.100%		Expires: 11/12/2015
Solution Control Lot:	13K125		Bottle No.: 1330

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:06S		
Control 1 EC	0.099%	08:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07S		
Control 2 EC	0.099%	08:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09S		
Control 3 EC	0.099%	08:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PFC #421 Badge No.: 421

Date: 02/16/2015