Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00705	Calib. Date:	09/30/2014	Calib. No.:	00029
Certification File No.:	00706	Cert. Date:	09/30/2014	Cert. No.:	00019
Linearity File No.:	00707	Lin. Date:	09/30/2014	Lin. No.:	00019
Solution File No.:	00742	Soln. Date:	03/17/2015	Soln. No.:	00161
Sequential File No.:	00742	File Date:	03/17/2015		

Calibrating Unit:	4	Model No.:	CU-34	Serial No.:	DDUF S3-0061
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:28S		
Control 1 EC	0.100%	06:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:30S		
Control 2 EC	0.100%	06:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	06:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:31S		
Control 3 EC	0.100%	06:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH. #421 Badge No.: 421

Date: 03/17/2015