

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00376 Calib. Date: 12/03/2010 Calib. No.: 00014
Certification File No.: 00377 Cert. Date: 12/03/2010 Cert. No.: 00010
Linearity File No.: 00378 Lin. Date: 12/03/2010 Lin. No.: 00010
Solution File No.: 00404 Soln. Date: 03/28/2011 Soln. No.: 00095
Sequential File No.: 00404 File Date: 03/28/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 0150

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:43S | | |
| Control 1 EC | 0.102% | 07:44S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.103% | 07:44S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.008% | 07:45S | | |
| Control 2 EC | 0.096% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.096% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:46S | | |
| Control 3 EC | 0.094% | 07:47S | 34.0°C | CONTROL TEST FAILED |
| Control 3 IR | 0.095% | 07:47S | 34.0°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 07:48S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 03/28/2011

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARNK-0074 |
| Location: | SOUTH BRUNSWICK POLICE | Calib. No.: | 00014 |
| Calibration File No.: | 00376 | Calib. Date: | 12/03/2010 |
| Certification File No.: | 00377 | Cert. Date: | 12/03/2010 |
| Linearity File No.: | 00378 | Lin. Date: | 12/03/2010 |
| Solution File No.: | 00404 | Soln. Date: | 03/28/2011 |
| Sequential File No.: | 00404 | File Date: | 03/28/2011 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0060 |
| Solution Control Lot: | 10L084 | Expires: | 12/16/2012 |
| | | Bottle No.: | 0150 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 07:43S | | |
| Control 1 EC | 0.102% | 07:44S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.103% | 07:44S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.008% | 07:45S | | |
| Control 2 EC | 0.096% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.096% | 07:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:46S | | |
| Control 3 EC | 0.094% | 07:47S | 34.0°C | CONTROL TEST FAILED |
| Control 3 IR | 0.095% | 07:47S | 34.0°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 07:48S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/28/2011