Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00743	Calib. Date:	03/19/2015	Calib. No.:	00030
Certification File No.:	00744	Cert. Date:	03/19/2015	Cert. No.:	00020
Linearity File No.:	00745	Lin. Date:	03/19/2015	Lin. No.:	00020
Solution File No.:	00759	Soln. Date:	05/13/2015	Soln. No.:	00164
Sequential File No.:	00759	File Date:	05/13/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0033

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:38D		
Control 1 EC	0.100%	11:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:39D		
Control 2 EC	0.100%	11:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:41D		
Control 3 EC	0.100%	11:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: Pfl. #421

Badge No.: 421

Date: 05/13/2015