Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00673	Calib. Date:	04/07/2014	Calib. No.:	00028
Certification File No.:	00674	Cert. Date:	04/07/2014	Cert. No.:	00018
Linearity File No.:	00675	Lin. Date:	04/07/2014	Lin. No.:	00018
Solution File No.:	00692	Soln. Date:	07/01/2014	Soln. No.:	00150
Sequential File No.:	00692	File Date:	07/01/2014		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0061
Control Solution %:	0.100%			Expires:	11/12/2015
Solution Control Lot:	13K125			Bottle No.:	1339

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:48D		
Control 1 EC	0.099%	19:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:50D		
Control 2 EC	0.098%	19:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:51D		
Control 3 EC	0.099%	19:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH. #421

Badge No.: 421

Date: 07/01/2014