

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00263 Calib. Date: 06/24/2009 Calib. No.: 00011  
Certification File No.: 00264 Cert. Date: 06/24/2009 Cert. No.: 00007  
Linearity File No.: 00265 Lin. Date: 06/24/2009 Lin. No.: 00007  
Solution File No.: 00270 Soln. Date: 07/22/2009 Soln. No.: 00069  
Sequential File No.: 00270 File Date: 07/22/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 12/01/2010  
Solution Control Lot: 08L061 Bottle No.: 1075

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:36D		
Control 1 EC	0.099%	12:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38D		
Control 2 EC	0.099%	12:38D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39D		
Control 3 EC	0.099%	12:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_



Badge No.: 258

Date: 07/22/2009