

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00411 Calib. Date: 05/04/2011 Calib. No.: 00017  
Certification File No.: 00412 Cert. Date: 05/04/2011 Cert. No.: 00011  
Linearity File No.: 00413 Lin. Date: 05/04/2011 Lin. No.: 00011  
Solution File No.: 00423 Soln. Date: 08/02/2011 Soln. No.: 00102  
Sequential File No.: 00423 File Date: 08/02/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 12/16/2012  
Solution Control Lot: 10L084 Bottle No.: 0138

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:36D		
Control 1 EC	0.099%	09:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 2 EC	0.099%	09:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 3 EC	0.099%	09:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 08/02/2011

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00411	Calib. Date:	05/04/2011
Certification File No.:	00412	Cert. Date:	05/04/2011
Linearity File No.:	00413	Lin. Date:	05/04/2011
Solution File No.:	00424	Soln. Date:	08/03/2011
Sequential File No.:	00424	File Date:	08/03/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0138

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:33D		
Control 1 EC	0.099%	13:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		
Control 2 EC	0.099%	13:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		
Control 3 EC	0.099%	13:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 08/03/2011