

# Calibrating Unit

## New Standard Solution Report

**Equipment**


Alcotest 7110 MKIII-C		Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00785	Calib. Date: 08/20/2015
Certification File No.:	00786	Cert. No.: 00031
Linearity File No.:	00787	Cert. Date: 08/20/2015
Solution File No.:	00819	Lin. No.: 00021
Sequential File No.:	00819	Lin. Date: 08/20/2015
		Soln. No.: 00173
		Soln. Date: 12/15/2015
		File Date: 12/15/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 0081

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:16S		
Control 1 EC	0.100%	12:16S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:16S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17S		
Control 2 EC	0.100%	12:18S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:18S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18S		
Control 3 EC	0.101%	12:19S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:19S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE	First Name: JESSE	MI: R
Signature: <u></u>	Badge No.: 421	Date: 12/15/2015