

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C		Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE		
Calibration File No.: 00531	Calib. Date: 10/22/2012	Calib. No.: 00025
Certification File No.: 00532	Cert. Date: 10/22/2012	Cert. No.: 00015
Linearity File No.: 00533	Lin. Date: 10/22/2012	Lin. No.: 00015
Solution File No.: 00542	Soln. Date: 12/17/2012	Soln. No.: 00127
Sequential File No.: 00542	File Date: 12/17/2012	
Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %: 0.100%		Expires: 06/28/2014
Solution Control Lot: 12F102		Bottle No.: 0402

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:05S		
Control 1 EC	0.100%	15:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		
Control 2 EC	0.100%	15:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08S		
Control 3 EC	0.100%	15:08S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_

Badge No.: 351

Date: 12/17/2012