

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00912 Calib. Date: 05/31/2017 Calib. No.: 00035
Certification File No.: 00913 Cert. Date: 05/31/2017 Cert. No.: 00025
Linearity File No.: 00914 Lin. Date: 05/31/2017 Lin. No.: 00025
Solution File No.: 00936 Soln. Date: 11/08/2017 Soln. No.: 00201
Sequential File No.: 00936 File Date: 11/08/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 03/20/2019
Solution Control Lot: 17110 Bottle No.: 0805

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:32S		
Control 1 EC	0.101%	15:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33S		
Control 2 EC	0.101%	15:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35S		
Control 3 EC	0.101%	15:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Pt. JML # 421

Badge No.: 421

Date: 11/08/2017