

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00940	Calib. Date: 11/24/2017
Certification File No.:	00941	Cert. No.: 00036
Linearity File No.:	00942	Cert. Date: 11/24/2017
Solution File No.:	00950	Lin. Date: 11/24/2017
Sequential File No.:	00950	Lin. No.: 00026
		Soln. Date: 01/19/2018
		Soln. No.: 00204
		File Date: 01/19/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0813

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:42S		
Control 1 EC	0.097%	15:42S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44S		
Control 2 EC	0.094%	15:44S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.101%	15:44S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:45S		
Control 3 EC	0.093%	15:46S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.101%	15:46S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:47S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 01/19/2018