Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00940	Calib. Date:	11/24/2017	Calib. No.:	00036
Certification File No.:	00941	Cert. Date:	11/24/2017	Cert. No.:	00026
Linearity File No.:	00942	Lin. Date:	11/24/2017	Lin. No.:	00026
Solution File No.:	00953	Soln. Date:	01/30/2018	Soln. No.:	00206
Sequential File No.:	00953	File Date:	01/30/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	03/20/2019
Solution Control Lot:	17110			Bottle No.:	0871

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:58S		
Control 1 EC	0.100%	14:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		
Control 2 EC	0.098%	15:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02S		
Control 3 EC	0.097%	15:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH #421 Badge No.: 421

Date: 01/30/2018