# Calibrating Unit
# New Standard Solution Report

| | | | | | |
|---|---|---|---|---|---|
| **Equipment** | Alcotest 7110 MKIII-C | | | Serial No.: | ARNK-0074 |
| Location: | SOUTH BRUNSWICK POLICE | | | | |
| Calibration File No.: | 00958 | Calib. Date: | 03/27/2018 | Calib. No.: | 00039 |
| Certification File No.: | 00959 | Cert. Date: | 03/27/2018 | Cert. No.: | 00027 |
| Linearity File No.: | 00960 | Lin. Date: | 03/27/2018 | Lin. No.: | 00027 |
| Solution File No.: | 00961 | Soln. Date: | 03/27/2018 | Soln. No.: | 00208 |
| Sequential File No.: | 00961 | File Date: | 03/27/2018 | | |

| | | | | | |
|---|---|---|---|---|---|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUF S3-0061 |
| Control Solution %: | 0.100% | | | Expires: | 03/20/2019 |
| Solution Control Lot: | 17110 | | | Bottle No.: | 0671 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|---|---|---|---|---|
| Ambient Air Blank | 0.000% | 10:21S | 11:21 D OK | |
| Control 1 EC | 0.101% | 10:22S | 33.8°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 10:22S | 33.8°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:22S | | |
| Control 2 EC | 0.101% | 10:23S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 10:23S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:23S | | |
| Control 3 EC | 0.102% | 10:24S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 10:24S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:24S | 11:24 D OK | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2-215 OK

**Changed By:**

Last Name: KLIMIK First Name: DAVID MI: W

Signature: TPI D. Kl #1040 Badge No.: 7040

Date: 03/27/2018