

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00958	Calib. Date: 03/27/2018
Certification File No.:	00959	Calib. No.: 00039
Linearity File No.:	00960	Cert. Date: 03/27/2018
Solution File No.:	00964	Cert. No.: 00027
Sequential File No.:	00964	Lin. Date: 03/27/2018
		Lin. No.: 00027
		Soln. Date: 04/25/2018
		Soln. No.: 00209
		File Date: 04/25/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0866

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:08D		
Control 1 EC	0.101%	08:09D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10D		
Control 2 EC	0.101%	08:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11D		
Control 3 EC	0.101%	08:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 04/25/2018