Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00958	Calib. Date:	03/27/2018	Calib. No.:	00039
Certification File No.:	00959	Cert. Date:	03/27/2018	Cert. No.:	00027
Linearity File No.:	00960	Lin. Date:	03/27/2018	Lin. No.:	00027
Solution File No.:	00976	Soln. Date:	08/08/2018	Soln. No.:	00213
Sequential File No.:	00976	File Date:	08/08/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0061
Control Solution %:	0.100%			Expires:	12/11/2019
Solution Control Lot:	17420			Bottle No.:	0943

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:44D		
Control 1 EC	0.099%	11:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 2 EC	0.099%	11:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 3 EC	0.100%	11:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: #421 Badge No.: 421

Date: 08/08/2018