Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	01053	Calib. Date:	07/10/2019	Calib. No.:	00042
Certification File No.:	01054	Cert. Date:	07/10/2019	Cert. No.:	00030
Linearity File No.:	01055	Lin. Date:	07/10/2019	Lin. No.:	00030
Solution File No.:	01059	Soln. Date:	08/07/2019	Soln. No.:	00228
Sequential File No.:	01059	File Date:	08/07/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1468

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:03D		
Control 1 EC	0.100%	08:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04D		
Control 2 EC	0.099%	08:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:05D		
Control 3 EC	0.099%	08:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH #421 Badge No.: 421

Date: 08/07/2019