

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01053	Calib. Date:	07/10/2019
Certification File No.:	01054	Cert. Date:	07/10/2019
Linearity File No.:	01055	Lin. Date:	07/10/2019
Solution File No.:	01077	Soln. Date:	09/25/2019
Sequential File No.:	01077	File Date:	09/25/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%		Serial No.:
Solution Control Lot:	19020		DDUF S3-0060
			Expires: 01/08/2021
			Bottle No.: 0259

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:06D		
Control 1 EC	0.100%	08:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		
Control 2 EC	0.100%	08:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09D		
Control 3 EC	0.100%	08:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #421

Badge No.: 421

Date: 09/25/2019