Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00424	Calib. Date:	04/16/2015	Calib. No.:	00031
Certification File No.:	00425	Cert. Date:	04/16/2015	Cert. No.:	00022
Linearity File No.:	00426	Lin. Date:	04/16/2015	Lin. No.:	00022
Solution File No.:	00430	Soln. Date:	07/09/2015	Soln. No.:	00159
Sequential File No.:	00430	File Date:	07/09/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0080
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:16D		
Control 1 EC	0.100%	08:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18D		
Control 2 EC	0.099%	08:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		
Control 3 EC	0.100%	08:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: #421 Badge No.: 421

Date: 07/09/2015