

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00221	Calib. Date: 06/24/2009
Certification File No.:	00222	Calib. No.: 00010
Linearity File No.:	00223	Cert. Date: 06/24/2009
Solution File No.:	00226	Cert. No.: 00008
Sequential File No.:	00226	Lin. Date: 06/24/2009
		Lin. No.: 00008
		Soln. Date: 08/20/2009
		Soln. No.: 00062
		File Date: 08/20/2009

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0060
Control Solution %: 0.100%		Expires: 12/01/2010
Solution Control Lot: 08L061		Bottle No.: 1055

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:37D		
Control 1 EC	0.101%	12:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38D		
Control 2 EC	0.101%	12:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39D		
Control 3 EC	0.100%	12:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 08/20/2009