

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00454	Calib. Date: 07/11/2016
Certification File No.:	00455	Calib. No.: 00034
Linearity File No.:	00456	Cert. Date: 07/11/2016
Solution File No.:	00462	Cert. No.: 00025
Sequential File No.:	00462	Lin. Date: 07/11/2016
		Lin. No.: 00025
		Soln. Date: 09/23/2016
		Soln. No.: 00177
		File Date: 09/23/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: 4
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1030

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:33D		
Control 1 EC	-.---%	09:33D	22.6°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	09:33D	22.6°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  # 421

Badge No.: 421

Date: 09/23/2016

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00454	Calib. No.: 00034
Certification File No.:	00455	Cert. No.: 00025
Linearity File No.:	00456	Lin. No.: 00025
Solution File No.:	00463	Soln. No.: 00178
Sequential File No.:	00463	File Date: 09/28/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: 4
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:05D		
Control 1 EC	0.098%	08:06D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	08:06D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	08:06D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 09/28/2016

