Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00433	Calib. Date:	08/20/2015	Calib. No.:	00032
Certification File No.:	00434	Cert. Date:	08/20/2015	Cert. No.:	00023
Linearity File No.:	00435	Lin. Date:	08/20/2015	Lin. No.:	00023
Solution File No.:	00438	Soln. Date:	10/20/2015	Soln. No.:	00163
Sequential File No.:	00438	File Date:	10/20/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0047

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:29D		
Control 1 EC	0.100%	07:29D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:29D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:30D		
Control 2 EC	0.099%	07:31D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:31D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:31D		
Control 3 EC	0.099%	07:32D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:32D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: M. #421 Badge No.: 421

Date: 10/20/2015