Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00524	Calib. Date:	03/27/2018	Calib. No.:	00038
Certification File No.:	00525	Cert. Date:	03/27/2018	Cert. No.:	00029
Linearity File No.:	00526	Lin. Date:	03/27/2018	Lin. No.:	00029
Solution File No.:	00527	Soln. Date:	03/27/2018	Soln. No.:	00201
Sequential File No.:	00527	File Date:	03/27/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	03/20/2019
Solution Control Lot:	17110			Bottle No.:	1061

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:44S	13:44 D DR	
Control 1 EC	0.100%	12:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46S		
Control 2 EC	0.099%	12:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 3 EC	0.100%	12:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49S	13:49 D DK	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSF P2-0394 DK

Changed By:

Last Name: KLIMIK First Name: DAVID MI: W

Signature: DAVID W. K #7040 Badge No.: 7040

Date: 03/27/2018