Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00560	Calib. Date:	07/10/2019	Calib. No.:	00041
Certification File No.:	00561	Cert. Date:	07/10/2019	Cert. No.:	00032
Linearity File No.:	00562	Lin. Date:	07/10/2019	Lin. No.:	00032
Solution File No.:	00564	Soln. Date:	08/07/2019	Soln. No.:	00225
Sequential File No.:	00564	File Date:	08/07/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0080
Control Solution %:	0.100%			Expires:	01/08/2021
Solution Control Lot:	19020			Bottle No.:	0343

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:39D		
Control 1 EC	0.101%	08:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:40D		
Control 2 EC	0.100%	08:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:41D		
Control 3 EC	0.100%	08:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: DH. #421 Badge No.: 421

Date: 08/07/2019