Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00560	Calib. Date:	07/10/2019	Calib. No.:	00041
Certification File No.:	00561	Cert. Date:	07/10/2019	Cert. No.:	00032
Linearity File No.:	00562	Lin. Date:	07/10/2019	Lin. No.:	00032
Solution File No.:	00565	Soln. Date:	09/04/2019	Soln. No.:	00226
Sequential File No.:	00565	File Date:	09/04/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0080
Control Solution %:	0.100%			Expires:	01/08/2021
Solution Control Lot:	19020			Bottle No.:	0276

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:59D		
Control 1 EC	0.101%	07:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		
Control 2 EC	0.099%	08:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01D		
Control 3 EC	0.099%	08:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PH. #421

Badge No.: 421

Date: 09/04/2019