Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE				
Calibration File No.:	00546	Calib. Date:	01/14/2019	Calib. No.:	00040
Certification File No.:	00547	Cert. Date:	01/14/2019	Cert. No.:	00031
Linearity File No.:	00548	Lin. Date:	01/14/2019	Lin. No.:	00031
Solution File No.:	00550	Soln. Date:	02/05/2019	Soln. No.:	00216
Sequential File No.:	00550	File Date:	02/05/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1123

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:32S		
Control 1 EC	0.101%	13:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33S		
Control 2 EC	0.099%	13:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35S		
Control 3 EC	0.100%	13:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: pH #421

Badge No.: 421

Date: 02/05/2019